6th Risk, Diagnosis and Treatment of Cardiovascular Disease in Women

# Where We Stand on Left Atrial Appendage Closure for Stroke Prevention in Atrial Fibrillation

Payam Safavi-Naeini, MD Abdi Rasekh, MD

#### ★ CME Credit

Presented at the 6th Annual Symposium on Risk, Diagnosis and Treatment of Cardiovascular Disease in Women; Houston, 9 January 2016.

**Section Editor:** Stephanie A. Coulter, MD

Key words: Atrial fibrillation/complications; cerebrovascular disorders/etiology; left atrial appendage/ligation; risk factors; stroke/ prevention & control; thromboembolism/prevention & control; warfarin/therapeutic use

From: Department of Cardiology (Drs. Rasekh and Safavi-Naeini), Texas Heart Institute; and Department of Medicine, Section of Cardiology (Dr. Rasekh), Baylor College of Medicine; Houston, Texas 77030

### Address for reprints:

Abdi Rasekh, MD, 6624 Fannin St., Houston, TX 77030

E-mail: arasekh@bcm.edu

© 2016 by the Texas Heart® Institute, Houston trial fibrillation (AF) is the most prevalent cardiac arrhythmia—found in 1% to 2% of the general population.¹ It is also a major cause of ischemic stroke. Patients with AF are 4 to 5 times more likely to have an ischemic stroke.² Their annual rate of stroke is 5%, which accounts for 15% of all strokes in the United States.².³ The left atrial appendage (LAA) is a trabeculated cul-de-sac that is responsible for more than 90% of thrombus formations in AF patients.⁴ Because of these findings, LAA closure devices have been developed as an alternative to oral anticoagulant (OAC) therapy.

The first generation of endocardial LAA occluders comprised 3 devices: the Percutaneous LAA Transcatheter Occlusion system (PLAATO) (ev3 Inc.; Plymouth, Minn), the WATCHMAN<sup>TM</sup> Left Atrial Appendage Closure Device (Boston Scientific Corporation; Natick, Mass), and the AMPLATZER<sup>TM</sup> Cardiac Plug (ACP) device (St. Jude Medical, Inc.; Minneapolis, Minn). PLAATO withdrew from the market in 2006, and the ACP is not yet available for commercial use in the U.S.<sup>5</sup>

The newer generation of endocardial LAA occluders comprises the WaveCrest® LAA Occlusion System (Coherex Medical, Inc.; Salt Lake City, Utah), the LAmbre™ Left Atrial Appendage Occluder (Lifetech Scientific [Shenzhen] Co., Ltd.; Shenzhen, PRC), and the Ultrasept LAA Closure Device (Cardia Inc.; Eagan, Minn), none of which is available for commercial use in the U.S. If you prefer to perform LAA closure through an epicardial approach, the LARIAT® Suture Delivery Device (Sentre-HEART, Inc.; Redwood City, Calif) and the Aegis device (Aegis Medical Innovations Inc.; Vancouver, Canada) will permit that. But the Aegis device is available for investigational use only. Inc.

## The Watchman Clinical Trial Experience

The WATCHMAN LAA Closure Technology for Embolic Protection in Patients with Atrial Fibrillation (PROTECT AF) clinical trial was designed to examine the safety and efficacy of the WATCHMAN device in patients with nonvalvular AF who were eligible for warfarin therapy and had a CHADS<sub>2</sub> stroke-risk score of 1 or greater.<sup>7</sup> From February 2005 through June 2008, 707 patients at 59 centers in the U.S. and Europe were enrolled in the PROTECT AF study. The patients were randomly assigned in a 2:1 ratio to percutaneous closure of the LAA and subsequent discontinuation of warfarin (n=463), or to warfarin treatment. All patients were monitored for 12 months.<sup>7</sup> The PROTECT AF trial showed that the WATCHMAN LAA Closure is noninferior to warfarin for the combined endpoint of stroke, systemic embolism, and death, and it is a safe and effective alternative to OAC therapy in decreasing the risk of AF-related stroke.<sup>78</sup> The PROTECT AF trial showed a high rate (8.7%) of implant device-related sequelae, including perforations and pericardial effusions, and had a procedural failure rate of around 9%. However, most of these events occurred during the early stages of the trial, and procedural failure rates declined as surgeons became more experienced with the operation.9

According to the PROTECT AF trial, the U.S. Food and Drug Administration (FDA) review panel required a 2nd trial to show that the lower procedural failure rate at the late stages of the study could be maintained across various hospitals and with new operators. In order to respond to the FDA panel's concerns, a 2nd trial, The Prospec-

tive Randomized EVAluation (PREVAIL), was designed, involving 407 patients randomly distributed in a 2:1 ratio to warfarin (as a control group) or to WATCHMAN LAA Closure. These 407 patients were warfarin-eligible and had a mean CHADS $_2$  score of  $2.6\pm1.10$  The Prevail study showed an overall 95% success rate for implanting the device (93.2% for new surgeons), and the overall 7-day serious procedure- or device-related complication rate was 4.4% (a 49% relative reduction rate) in comparison with Protect AF, which had an 8.7% device-related complication rate.

The PROTECT AF/PREVAIL meta-analysis (2,406 patients and 5,931 years of patient follow-up) showed that rates of hemorrhagic stroke, nonprocedural bleeding, and cardiovascular and unexplained deaths were decreased among patients who had LAA closure with use of the WATCHMAN, versus patients who were on long-term OAC therapy.<sup>11</sup> It bears mention, however, that if periprocedural complications were included, there would be no significant differences in rates of all-cause death or major bleeding.<sup>12</sup>

#### **LARIAT Studies**

The LARIAT is a minimally invasive procedure to tie off the LAA and remove the main source of AF-related stroke for patients who are at high risk of stroke and are not suitable candidates for OAC therapy. A single-center, nonrandomized study (PLACE II) showed about a 96% incidence of effective LAA closure (85 of 89 procedures). Three adverse events (3.3%) were reported, in-

cluding bleeding (2 pericardial and 1 transseptal). The study investigators determined that the LARIAT procedure can effectively close the LAA, with acceptable low-access sequelae and periprocedural adverse events.<sup>13</sup>

To date, the LARIAT has not been involved in any prospective randomized trials, but published clinical results for single-center studies (together with a recent multicenter study) provide insight into the potential of the LARIAT as a method of LAA closure, as well as insight into the potential for protection against stroke (Table I).13-17 These study results indicate consistent trends of efficacy and safety for LAA closure, which warrant further evaluation of the LARIAT in a multicenter prospective randomized trial. Sievert and colleagues<sup>17</sup> conducted a 5-center study that evaluated the use of LARIAT for patients who had contraindications to OAC therapy, such as histories of bleeding, stroke/transient ischemic attack, or cerebral aneurysm. There were 6 deaths during this study's 2-year follow-up period, one of them procedure-related because of a pulmonary embolus one day after the procedure. Sievert and colleagues' study provides promising insight into the potential of LARIAT's percutaneous, nonimplant approach in high-risk patients for whom there are no other options.<sup>17</sup>

# Impact of LARIAT LAA Exclusion on Atrial Fibrillation Burden

Lakkireddy and colleagues<sup>18</sup> compared the AF burden over a period of 3 months, before and after LAA ligation. Among the 18 patients in that study, the AF bur-

TABLE I. Efficacy and Sequelae of the Lariat Procedure<sup>14</sup>

Study	LARIAT PLACE II Bartus K, et al. <sup>13</sup> (2013)	LARIAT Massumi A, et al. <sup>15</sup> (2013)	LARIAT Stone D, et al. <sup>16</sup> (2015)	LARIAT No OAC Sievert H, et al. <sup>17</sup> (2015)	Lariat Cumulative
Patients (n)	89	21	27	143	280
Intent to treat	85 (96)	20 (95)	25 (93)	139 (97)	269 (96)
Procedural closure among intent-to-treat population	82 (96)	19 (95)	25 (100)	138 (99)	264 (98)
>60-day closure among patients who had follow-up TEE	81 (95)	16 (94)	22 (100)	126 (91)	245 (91)
CHADS <sub>2</sub> score	$1.9 \pm 0.95$	$3.2 \pm 1.2$	$3.5 \pm 1.4$	$2.4 \pm 1.2$	$2.6\pm1.2$
Sequelae					
Device-related	0	0	0	0	0
Access-related	3 (3.4)	1 (4.8)	1 (3.7)	3 (2.1)	8 (2.9)
All-cause death	2 (2.2)	1 (4.8)	0	6 (4.2)	4 (1.4)
All-cause stroke	2 (2.2)	0	1 (3.7)	4 (2.8)	7 (2.5)
Major bleeding	0	0	1 (3.7)	2 (1.4)	3 (1.1)
Pericardial/pleural effusion	1 (1.1)	3 (14.3)	2 (7.4)	1 (0.7)	7 (2.5)

OAC = oral anticoagulation; TEE = transesophageal echocardiography

Data are presented as mean  $\pm$  SD or as number and percentage.

Adapted with permission from J Tehran Heart Cent 2015;10(2):69-73.14

den decreased from a baseline of 81% to 47% after ligation (P < 0.01). Afzal and associates recorded the AF burden before LAA exclusion (baseline), and twice later, at 3 and 12 months after exclusion. According to this study, the AF burden at the 3-month follow-up evaluation was  $42\% \pm 34\%$ , significantly lower than that at baseline ( $76\% \pm 33\%$ ; P < 0.0001); and the reduction of AF burden was fairly well sustained at 12 months ( $59\% \pm 26\%$ ; P < 0.001).

#### The aMAZE Trial

The LAA Ligation Adjunctive to PVI for Persistent or Long-Standing Persistent Atrial Fibrillation (aMAZE) Trial to Evaluate LARIAT Ligation of the Left Atrial Appendage is a prospective, multicenter, randomized (2:1) controlled study.<sup>20</sup> This trial was designed to evaluate the safety and effectiveness of the LARIAT, which percutaneously isolates and ligates the LAA from the left atrium as an adjunct to planned catheter ablation of the pulmonary vein in isolation, in the treatment of patients who manifest symptomatic persistent or long-standing persistent AF. This study will be conducted in 2 stages: in Stage 1 (the Limited Early Stage), there will be as many as 175 patients at up to 15 sites; and in Stage 2 (the Pivotal Stage), there will be as many as 600 patients at up to 50 sites. All patients from both stages will be included in the primary analysis.

#### Conclusion

The WATCHMAN is the only LAA closure device approved by the FDA for LAA exclusion in the U.S. Moreover, there are still no approved devices for patients who have AF and a contraindication to OAC therapy. Several retrospective multicenter studies of experiences with the LARIAT device have confirmed high acute procedural success rates. The role of LAA ligation in reducing AF burdens in patients who display persistent AF must be further evaluated in a multicenter prospective study. The role of LARIAT in the treatment of a targeted population with persistent AF is currently under investigation. Co

#### References

- Malladi V, Naeini PS, Razavi M, Collard CD, Anton JM, Tolpin DA. Endovascular ablation of atrial fibrillation. Anesthesiology 2014;120 (6):1513-9.
- Wolf PA, Abbott RD, Kannel WB. Atrial fibrillation as an independent risk factor for stroke: the Framingham Study. Stroke 1991;22(8):983-8.
- Risk factors for stroke and efficacy of antithrombotic therapy in atrial fibrillation. Analysis of pooled data from five randomized controlled trials [published erratum appears in Arch Intern Med 1994;154(19):2254]. Arch Intern Med 1994;154 (13):1449-57.
- Cardona L, Galrinho A, Branco L, Leal A, Fiarresga A, Sousa L, Ferreira RC. Thrombus formation on a left atrial appendage closure device. Circulation 2011;124(14):1595-6.

- De Backer O, Arnous S, Ihlemann N, Vejlstrup N, Jorgensen E, Pehrson S, et al. Percutaneous left atrial appendage occlusion for stroke prevention in atrial fibrillation: an update. Open Heart 2014;1(1):e000020.
- Romero J, Perez IE, Krumerman A, Garcia MJ, Lucariello RJ. Left atrial appendage closure devices. Clin Med Insights Cardiol 2014;8:45-52.
- Holmes DR, Reddy VY, Turi ZG, Doshi SK, Sievert H, Buchbinder M, et al. Percutaneous closure of the left atrial appendage versus warfarin therapy for prevention of stroke in patients with atrial fibrillation: a randomised non-inferiority trial [published erratum appears in Lancet 2009;374(9701):1596]. Lancet 2009;374(9689):534-42.
- Reddy VY, Doshi SK, Sievert H, Buchbinder M, Neuzil P, Huber K, et al. Percutaneous left atrial appendage closure for stroke prophylaxis in patients with atrial fibrillation: 2.3-year follow-up of the PROTECT AF (Watchman Left Atrial Appendage System for Embolic Protection in Patients with Atrial Fibrillation) Trial. Circulation 2013;127(6):720-9.
- Reddy VY, Holmes D, Doshi SK, Neuzil P, Kar S. Safety of percutaneous left atrial appendage closure: results from the Watchman Left Atrial Appendage System for Embolic Protection in Patients with AF (PROTECT AF) clinical trial and the Continued Access Registry. Circulation 2011;123(4):417-24.
- Holmes DR Jr, Kar S, Price MJ, Whisenant B, Sievert H, Doshi SK, et al. Prospective randomized evaluation of the Watchman Left Atrial Appendage Closure device in patients with atrial fibrillation versus long-term warfarin therapy: the PREVAIL Trial [published erratum appears in J Am Coll Cardiol 2014;64(11):1186]. J Am Coll Cardiol 2014;64(1):1-12.
- 11. Holmes DR Jr, Doshi SK, Kar S, Price MJ, Sanchez JM, Sievert H, et al. Left atrial appendage closure as an alternative to warfarin for stroke prevention in atrial fibrillation: a patient-level meta-analysis. J Am Coll Cardiol 2015;65(24):2614-23.
- Sohaib SM, Fox KF. A meta-analysis of left atrial appendage closure for stroke prevention in atrial fibrillation-adding to the debate but elements remain unresolved. J Thorac Dis 2015;7 (8):F.226-9.
- Bartus K, Han FT, Bednarek J, Myc J, Kapelak B, Sadowski J, et al. Percutaneous left atrial appendage suture ligation using the LARIAT device in patients with atrial fibrillation: initial clinical experience. J Am Coll Cardiol 2013;62(2):108-18.
- Safavi-Naeini P, Razavi M, Saeed M, Rasekh A, Massumi A. A review of the LARIAT Suture Delivery Device for left atrial appendage closure. J Tehran Heart Cent 2015;10(2):69-73.
- Massumi A, Chelu MG, Nazeri A, May SA, Afshar-Kharaghan H, Saeed M, et al. Initial experience with a novel percutaneous left atrial appendage exclusion device in patients with atrial fibrillation, increased stroke risk, and contraindications to anticoagulation. Am J Cardiol 2013;111(6):869-73.
- Stone D, Byrne T, Pershad A. Early results with the LARIAT device for left atrial appendage exclusion in patients with atrial fibrillation at high risk for stroke and anticoagulation. Catheter Cardiovasc Interv 2015;86(1):121-7.
- 17. Sievert H, Rasekh A, Bartus K, Morelli RL, Fang Q, Kuropka J, et al. Left atrial appendage ligation in nonvalvular atrial fibrillation patients at high risk for embolic events with ineligibility for oral anticoagulation. Initial report of clinical outcomes. JACEP 2015;1(6):465-74. Available from: http://electrophysiology.onlinejacc.org/article.aspx?articleid=2473396 [cited 2016 Jun 16].
- Lakkireddy DR, Earnest M, Janga P, Reddy M, Vallakati A, Nath J, et al. Effect of endoepicardial percutaneous left atrial appendage ligation (LARIAT) on arrhythmia burden in patients with atrial fibrillation [abstract]. J Am Coll Cardiol 2013;61(10\_S). Available from: http://Content.online

- jacc.org/article.aspx?articleid=1664532&resultClick=3 [cited 2016 Jun 16].
- 19. Afzal MR, Kanmanthareddy A, Earnest M, Reddy M, Atkins D, Bommana S, et al. Impact of left atrial appendage exclusion using an epicardial ligation system (LARIAT) on atrial fibrillation burden in patients with cardiac implantable electronic devices. Heart Rhythm 2015;12(1):52-9.
- aMAZE Study: LAA Ligation Adjunctive to PVI for Persistent or Longstanding Persistent Atrial Fibrillation (aMAZE).
  Available from: https://clinicaltrials.gov/ct2/show/NCT02513797 [2016 Apr 28; cited 2016 Jun 16].