

# Progress in Endovascular Aortic Repair for Women

David Kuten, MD  
Zvonimir Krajcer, MD

## ★ CME Credit

Presented at the 7th Annual  
Women's Heart & Vascular  
Symposium, Texas Heart  
Institute, Houston,  
10 December 2016.

**Section Editor:**  
Stephanie A. Coulter, MD

**Key words:** Aortic aneurysm, abdominal/epidemiology; blood vessel prosthesis implantation/instrumentation; delivery of health care; endovascular procedures/instrumentation; female; prevalence; prosthesis design; sex factors

**From:** Department of  
Cardiology, Texas Heart  
Institute and CHI St. Luke's  
Health—Baylor St. Luke's  
Medical Center, Houston,  
Texas 77030

**Address for reprints:**  
Zvonimir Krajcer, MD,  
6624 Fannin St., Suite 2780,  
Houston, TX 77030

**E-mail:** zvonkomd@aol.com

© 2017 by the Texas Heart®  
Institute, Houston

**T**he diagnosed prevalence of abdominal aortic aneurysm (AAA) in women is approximately 20% of that in men.<sup>1-3</sup> Therefore, women are underrepresented in epidemiologic studies and trials in which screening and therapeutic methods are evaluated. Consequently, central issues related to AAA in women have not been fully investigated.

First, screening recommendations provide inadequate guidance for women. The United States Preventive Services Task Force has concluded that insufficient evidence exists for screening women with a history of smoking and has advised against screening women who have never smoked.<sup>4</sup> The only prospective studies of ultrasound screening in women yielded no difference in mortality rates at 10 years.<sup>3,5</sup> Subsequent efforts to identify women at high risk of AAA have been more promising.<sup>6,7</sup>

Female sex is associated with a more rapid aneurysmal growth rate and earlier rupture.<sup>8,9</sup> These findings suggest that women might benefit more from earlier AAA repair than do men. In addition, AAAs in women are more likely to have complex features.<sup>10</sup> The median infrarenal aortic neck length in women is 12 mm, compared with 16 mm in men. The degree of neck angulation is greater in women (45° vs 36°), and iliac vessel caliber is smaller (11 vs 14 mm). Thus, approximately 60% of women are not eligible for endovascular aortic repair (EVAR) with use of conventional devices.<sup>11</sup>

Recent advances in device design, including smaller delivery systems and modifications to accommodate shorter neck lengths and greater angulation, can be expected to increase EVAR eligibility in women. For instance, the Endurant® II AAA Stent Graft System (Medtronic, Inc.; Minneapolis, Minn) and the TriVascular Ovation Prime® device (TriVascular, an Endologix company; Santa Rosa, Calif) can accommodate neck lengths as short as 10 mm and 7 mm, respectively. The Zenith® Fenestrated AAA Endovascular Graft (Cook Medical Inc.; Bloomington, Ind), which is uniquely manufactured to conform to patients' aortic anatomy, can be used in necks as short as 4 mm. Extreme neck angulation is accommodated by the Aorfix™ Endovascular Stent Graft (Lombard Medical, Inc.; Irvine, Calif). These devices can also be used in patients who have iliac vessel diameters as narrow as 8 mm and as wide as 25 mm. The TriVascular Ovation® provides the lowest-profile delivery system at 14F.

The ongoing LIFE Study seeks to show benefits associated with the low-profile TriVascular Ovation Prime device in conjunction with the Fast-Track EVAR Protocol, which includes percutaneous access, conscious sedation, and next-day discharge from the hospital.<sup>12</sup> Thirty-day results include low rates of adverse events, high procedural success rates, freedom from endoleak, and hospital readmission rates 5 times lower than those in contemporary EVAR reports (1.6% vs 8.2%).<sup>13</sup> Patients assigned to the fast-track approach showed a trend toward more greatly improved quality of life. Whether the predicted benefits of the Ovation Prime device will specifically translate into improved outcomes in female patients will be determined in the forthcoming LUCY Study.<sup>14</sup>

The underrepresentation of women in clinical trials related to AAA has resulted in a lack of guidance for patients and physicians in the central issues of natural history, screening, and treatment threshold. Technical advances in stent-grafts, which are now lower-profile and more compatible with complex aneurysmal features, are likely to increase eligibility for EVAR in women. In addition to procedural advances, future study should focus on improved screening algorithms and defining the optimal treatment threshold in women.

## References

1. DeRubertis BG, Trocciola SM, Ryer EJ, Pieracci FM, McKinsey JF, Faries PL, Kent KC. Abdominal aortic aneurysm in women: prevalence, risk factors, and implications for screening. *J Vasc Surg* 2007;46(4):630-5.
2. Lederle FA, Johnson GR, Wilson SE; Aneurysm Detection and Management Veterans Affairs Cooperative Study. Abdominal aortic aneurysm in women. *J Vasc Surg* 2001;34(1):122-6.
3. Scott RA, Bridgewater SG, Ashton HA. Randomized clinical trial of screening for abdominal aortic aneurysm in women. *Br J Surg* 2002;89(3):283-5.
4. Final Recommendation Statement: Abdominal aortic aneurysm: screening. U.S. Preventive Services Task Force. June 2014. Available at: <https://www.uspreventiveservicestaskforce.org/Page/Document/RecommendationStatementFinal/abdominal-aortic-aneurysm-screening> [2014 Jun].
5. Norman PE, Jamrozik K, Lawrence-Brown MM, Le MT, Spencer CA, Tuohy RJ, et al. Population based randomised controlled trial on impact of screening on mortality from abdominal aortic aneurysm [published erratum appears in *BMJ* 2005;330(7491):596]. *BMJ* 2004;329(7477):1259.
6. Chabok M, Nicolaides A, Aslam M, Farahmandfar M, Humphries K, Kermani NZ, et al. Risk factors associated with increased prevalence of abdominal aortic aneurysm in women. *Br J Surg* 2016;103(9):1132-8.
7. Kent KC, Zwolak RM, Egorova NN, Riles TS, Manganaro A, Moskowitz AJ, et al. Analysis of risk factors for abdominal aortic aneurysm in a cohort of more than 3 million individuals. *J Vasc Surg* 2010;52(3):539-48.
8. Brown LC, Powell JT. Risk factors for aneurysm rupture in patients kept under ultrasound surveillance. UK Small Aneurysm Trial Participants. *Ann Surg* 1999;230(3):289-97.
9. Solberg S, Singh K, Wilsgaard T, Jacobsen BK. Increased growth rate of abdominal aortic aneurysms in women. The Tromsø study. *Eur J Vasc Endovasc Surg* 2005;29(2):145-9.
10. Sweet MP, Fillinger MF, Morrison TM, Abel DB. The influence of gender and aortic aneurysm size on eligibility for endovascular abdominal aortic aneurysm repair [published erratum appears in *J Vasc Surg* 2012;55(1):310]. *J Vasc Surg* 2011;54(4):931-7.
11. Morrison TM, Fillinger MF, Meyer CA, Abel DB, Yan XS. Gender disparities in endovascular treatment options for infrarenal abdominal aortic aneurysms. Available at: <https://wayback.archive-it.org/7993/20161024042327/http://www.fda.gov/downloads/MedicalDevices/NewsEvents/WorkshopsConferences/UCM359044.pdf> [2013 Jun 25; cited 2017 Oct 12].
12. Krajcer Z, Ramaiah VG, Huetter M, Miller LE. Fast-track endovascular aortic repair: interim report from the prospective LIFE registry. *Catheter Cardiovasc Interv* 2016;88(7):1118-23.
13. Gupta PK, Fernandes-Taylor S, Ramanan B, Engelbert TL, Kent KC. Unplanned readmissions after vascular surgery. *J Vasc Surg* 2014;59(2):473-82.
14. LUCY Study: TriVascular evaluation of females who are underrepresented candidates for abdominal aortic aneurysm repair (LUCY). Available at: <https://clinicaltrials.gov/ct2/show/NCT02479191> [2015 Jun 24; updated 2017 May 3].