

Conduction System Pacing for Heart Failure

Nathan R. Smith, MD¹; Patrick Lynch, MD¹; Mihail G. Chelu, MD, PhD^{2,3,4}

¹Department of Internal Medicine, Baylor College of Medicine, Houston, Texas

²Department of Medicine, Section of Cardiology, Baylor College of Medicine, Houston, Texas

³Cardiovascular Research Institute, Baylor College of Medicine, Houston, Texas

⁴Department of Cardiology, The Texas Heart Institute, Houston, Texas



Keywords: Cardiac conduction system disease; cardiac pacing, artificial; cardiac resynchronization therapy; heart failure

Background

Biventricular pacing (BIVP) traditionally delivers cardiac resynchronization therapy (CRT) by using pacing leads in the right ventricle and the coronary sinus.

Cardiac resynchronization therapy is the heart failure (HF) therapy “that simultaneously improves cardiac function and functional capacity, reduces hospitalization, and prolongs survival”¹ in patients with HF with a reduced ejection fraction and a wide QRS complex.²⁻¹⁶ Solid randomized clinical trial (RCT) efficacy and safety data in more than 8,500 patients with biventricular devices^{17,18} have established CRT as the standard therapy in this category of patients. Approximately 20% to 40% of patients, however, do not respond to CRT via BIVP depending on the measure used.¹⁷ Up to 7% of biventricular pacemaker implants are furthermore unsuccessful as a result of difficulties encountered while implanting the left ventricular coronary sinus lead.⁹

For patients with a left ventricular ejection fraction of 36% to 50%, the BLOCK HF trial showed the superiority of BIVP to right ventricular pacing⁹ for a composite outcome of all-cause mortality, hospitalization as a result of HF, and an increase of more than 15% in the left ventricular end-systolic volume index. Because of higher costs, BIVP is not used often as a first-line therapy over right ventricular pacing outside of the United States.

Recent Developments

Because of the limitations of BIVP, conduction system pacing (CSP) at the level of His bundle¹⁹ or the left bundle branch²⁰ has emerged as an alternative physiologic pacing treatment that preserves or restores left ventricular electrical and mechanical synchrony. Left bundle branch area pacing has become the dominant approach because of its higher success rate, lower and more stable pacing thresholds, and its correction of the left bundle branch block below the level of the His bundle compared with His bundle pacing.²¹

There are only 7 randomized clinical trials comparing His bundle pacing and left bundle branch area pacing with BIVP.²¹ All of the trials are small, of short duration, and were not powered for major clinical end points such as HF hospitalizations and death. His bundle pacing and left bundle branch area pacing are at least equal or superior to BIVP in terms of surrogate parameters for electrical and mechanical synchrony.

There are substantially more data from large registries and retrospective comparative studies suggesting that His bundle pacing and left bundle branch area pacing may be superior to BIVP in terms of hard outcomes and possibly safety. An analysis of the safety of left bundle branch area pacing from the Multicentre European Left Bundle Branch

Citation: Smith NR, Lynch P, Chelu MG. Conduction system pacing for heart failure. *Tex Heart Inst J.* 2024;51(2):e248469.

doi:10.14503/THIJ-24-8469

Corresponding author: Mihail G. Chelu, 7200 Cambridge St, Suite A6.137, MS: BCM621, Houston, TX 77030 (mihail.chelu@bcm.edu)

Area Pacing Outcomes Study (MELOS) demonstrated higher success rates among early European adopters of left bundle branch area pacing when the procedure was performed for bradycardia (92.4%) and HF (82.2%).²² This study reported the highest complication rate (8.3%), though a notable majority of the complications were clinically insignificant septal perforations.

The largest retrospective case-control study, I-CLAS, suggested that left bundle branch area pacing outperformed BIVP and was associated with reduction in time to death and HF hospitalizations.²¹ Similar results were reported by 2 additional retrospective studies.^{23,24} Left bundle branch area pacing was also associated with a lower time to onset of both new-onset atrial fibrillation and ventricular arrhythmias, even in those patients with no history of ventricular arrhythmias who were naive to antiarrhythmic therapy.²⁵ A meta-analysis of 4 randomized and 17 observational studies showed CSP was associated with a significant reduction in all-cause mortality and HF hospitalizations compared with BIVP for CRT.²⁶

Future Directions

Despite the lack of large RCTs on these therapies, His bundle pacing and left bundle branch area pacing have been included in the most recent pacing guidelines for the avoidance and mitigation of HF as Class of Recommendation grades 2a and 2b, with similar indications to BIVP.²² Current Heart Rhythm Society, Asia Pacific Heart Rhythm Society, and Latin American Heart Rhythm Society guidelines recommend CSP as an alternative to traditional BIVP when effective CRT cannot be achieved (Class of Recommendation 2a).²² There are, however, multiple ongoing moderate to large RCTs that will fill gaps in clinical knowledge (Table I). Physiologic pacing for HF with a left ventricular ejection fraction of less than 50% remains a dynamic field

Abbreviations

BIVP, biventricular pacing
 CRT, cardiac resynchronization therapy
 CSP, conduction system pacing
 HF, heart failure
 RCT, randomized clinical trial

with multiple ongoing RCTs that will determine the relative benefits and safety of different pacing modalities. There is also an acute need to develop sheaths, leads, devices, and algorithms to improve and optimize the success rate of CSP.

Article Information

Published: 13 December 2024

Open Access: © 2024 The Authors. Published by The Texas Heart Institute®. This is an Open Access article under the terms of the Creative Commons Attribution-NonCommercial License (CC BY-NC, <https://creativecommons.org/licenses/by-nc/4.0/>), which permits use and distribution in any medium, provided the original work is properly cited, and the use is noncommercial.

Author Contributions: Nathan Smith assisted in the conceptualization of the manuscript; he wrote the initial draft of this paper, revised it, and incorporated revisions into the writing of subsequent drafts. Patrick Lynch assisted in the conceptualization of the manuscript and contributed to revision of the initial and subsequent drafts. Mihail Chelu conceived of the manuscript and contributed to the revisions of the initial and subsequent drafts.

Conflict of Interest Disclosure: None.

Funding/Support: Dr Chelu received research support from the Patient-Centered Outcomes Research Institute, National Institutes of Health, Impulse Dynamics, and Abbott. He received speaker's fees from Impulse Dynamics.

Section Editor: Mohammad Saeed, MD.

Meeting Presentation: Presented at: The Ali Massumi Cardiac Arrhythmia Symposium; February 24, 2024; Houston, TX.

TABLE I. Major Clinical Trials, by Left Ventricular Ejection Fraction Subpopulation, Comparing BVP, His Bundle Pacing, and Left Bundle Branch Area Pacing for CRT and Their Primary End Points

Left ventricular ejection fraction $\leq 35\%$	Left ventricular ejection fraction 36%-49%	Left ventricular ejection fraction $\geq 50\%$
CSP-SYNC [NCT05155865] (CSP vs BIVP) (N = 60) End points: left ventricular ejection fraction, left ventricular volume, New York Heart Association classification, 6-minute walk test, quality of life	His-PACE [NCT04672408] (His bundle pacing vs right ventricular pacing) (N = 50) End point: left ventricular ejection fraction	
HIS-alt_2 [NCT04409119] (CSP vs BIVP) (N = 125) End points: left ventricular end-systolic volume, QRS duration	HIS-PrEF [NCT04529577] (His bundle pacing vs right ventricular pacing) (N = 40) End point: left ventricular ejection fraction	
HIS-CRT [NCT05265520] (His bundle pacing vs BIVP) (N = 120) End point: left ventricular ejection fraction	LEAP [NCT04595487] (Left ventricular septal pacing vs right ventricular pacing) (N = 470) End points: death, HF hospitalizations, left ventricular ejection fraction decrease by 10%	
His-SYNC [NCT02700425] (His bundle pacing vs BIVP) (N = 41) End point: QRS duration	LEFT HF [NCT05015660] (Left bundle branch area pacing vs right ventricular pacing) (N = 1,280) End points: cardiovascular death, HF events	
LeCaRT [NCT05365568] (Left bundle branch area pacing vs BIVP) (N = 170) End points: composite death, HF hospitalizations, worsening HF, implant failure, cardiac implantable electrical device re-intervention	OptimPacing [NCT04624763] (Left bundle branch area pacing vs right ventricular pacing) (N = 683) End points: death, HF hospitalizations, pacemaker-induced cardiomyopathy	
LIT-HF [NCT05572957] (CSP vs guideline-directed medical therapy) (N = 50) End point: left ventricular ejection fraction	PROTECT-HF [NCT05815745] (His bundle pacing vs right ventricular pacing) (N = 2,600) End points: cardiovascular death, HF events	
Left-Bundle CRT [NCT05434962] (Left bundle branch area pacing vs BIVP) (N = 176) End point: CRT response	REINVENT-CRT [NCT05652218] (Left bundle branch area pacing vs BIVP) (N = 20) End point: myocardial perfusion imaging	
PhysioSync-HF [NCT05572736] (CSP vs BIVP) (N = 304) End points: death, HF hospitalizations, left ventricular ejection fraction		
CONSYST-CRT [NCT05187611] (CSP vs BIVP) (N = 130) End points: composite death, cardiac transplant, HF hospitalizations, and left ventricular ejection fraction		LBB Pacing Versus Conventional Pacing in Atrioventricular Block [NCT05722379] (Left bundle branch area pacing vs right ventricular pacing) (N = 27) End point: global work efficiency
HIPPOS [NCT05491655] (His bundle pacing vs backup right ventricular pacing) (N = 34) End point: left ventricular ejection fraction		RHYS PAVB [NCT05214365] (His bundle pacing vs right ventricular pacing) (N = 200) End point: pacemaker-induced cardiomyopathy
LBBAP-AFHF [NCT05549544] (Left bundle branch area pacing vs BIVP) (N = 60) End point: left ventricular ejection fraction		Vanguard [NCT05015660] (Left bundle branch area pacing vs right ventricular pacing) (N = 100) End points: death, HF events, left ventricular end-systolic volume, changes in lead parameters, quality of life; and safety
LBB Pacing in Patients with Cardiac Dysfunction and AV Block [NCT05553626] (Left bundle branch area pacing vs BIVP) (N = 160) End point: left ventricular ejection fraction		
Left vs Left [NCT05650658] (CSP vs BIVP) (N = 2,136) End points: death, HF hospitalizations, quality of life		

BIVP, biventricular pacing; CRT, cardiac resynchronization therapy; CSP, conduction system pacing; HF, heart failure.

References

- Auricchio A, Prinzen FW. Enhancing response in the cardiac resynchronization therapy patient: the 3B perspective-bench, bits, and bedside. *JACC Clin Electrophysiol.* 2017;3(11):1203-1219. doi:10.1016/j.jacep.2017.08.005
- Cazeau S, Leclercq C, Lavergne T, et al; Multisite Stimulation in Cardiomyopathies (MUSTIC) Study Investigators. Effects of multisite biventricular pacing in patients with heart failure and intraventricular conduction delay. *N Engl J Med.* 2001;344(12):873-880. doi:10.1056/NEJM200103223441202
- Auricchio A, Stellbrink C, Sack S, et al; Pacing Therapies in Congestive Heart Failure (PATH-CHF) Study Group. Long-term clinical effect of hemodynamically optimized cardiac resynchronization therapy in patients with heart failure and ventricular conduction delay. *J Am Coll Cardiol.* 2002;39(12):2026-2033. doi:10.1016/s0735-1097(02)01895-8
- Abraham WT, Fisher WG, Smith AL, et al; MIRACLE Study Group. Multicenter InSync Randomized Clinical Evaluation. Cardiac resynchronization in chronic heart failure. *N Engl J Med.* 2002;346(24):1845-1853. doi:10.1056/NEJMoa013168
- Higgins SL, Hummel JD, Niazi IK, et al. Cardiac resynchronization therapy for the treatment of heart failure in patients with intraventricular conduction delay and malignant ventricular tachyarrhythmias. *J Am Coll Cardiol.* 2003;42(8):1454-1459. doi:10.1016/s0735-1097(03)01042-8
- Young JB, Abraham WT, Smith AL, et al; Multicenter InSync ICD Randomized Clinical Evaluation (MIRACLE ICD) Trial Investigators. Combined cardiac resynchronization and implantable cardioversion defibrillation in advanced chronic heart failure: the MIRACLE ICD Trial. *JAMA.* 2003;289(20):2685-2694. doi:10.1001/jama.289.20.2685
- Sutton MGSJ, Plappert T, Hilpisch KE, Abraham WT, Hayes DL, Chinchoy E. Sustained reverse left ventricular structural remodeling with cardiac resynchronization at one year is a function of etiology: quantitative Doppler echocardiographic evidence from the Multicenter InSync Randomized Clinical Evaluation (MIRACLE). *Circulation.* 2006;113(2):266-272. doi:10.1161/CIRCULATIONAHA.104.520817
- Cleland JGF, Daubert JC, Erdmann E, et al; Cardiac Resynchronization-Heart Failure (CARE-HF) Study Investigators. The effect of cardiac resynchronization on morbidity and mortality in heart failure. *N Engl J Med.* 2005;352(15):1539-1549. doi:10.1056/NEJMoa050496
- Curtis AB, Worley SJ, Adamson PB, et al; Biventricular versus Right Ventricular Pacing in Heart Failure Patients with Atrioventricular Block (BLOCK HF) Trial Investigators. Biventricular pacing for atrioventricular block and systolic dysfunction. *N Engl J Med.* 2013;368(17):1585-1593. doi:10.1056/NEJMoa1210356
- Curtis AB, Worley SJ, Chung ES, Li P, Christman SA, Sutton MSJ. Improvement in clinical outcomes with biventricular versus right ventricular pacing: the BLOCK HF Study. *J Am Coll Cardiol.* 2016;67(18):2148-2157. doi:10.1016/j.jacc.2016.02.051
- Nagy KV, Merkely B, Rosero S, et al. Quality of life predicting long-term outcomes in cardiac resynchronization therapy patients. *Europace.* 2019;21(12):1865-1875. doi:10.1093/europace/euz262
- Bristow MR, Saxon LA, Boehmer J, et al; Comparison of Medical Therapy, Pacing, and Defibrillation in Heart Failure (COMPANION) Investigators. Cardiac-resynchronization therapy with or without an implantable defibrillator in advanced chronic heart failure. *N Engl J Med.* 2004;350(21):2140-2150. doi:10.1056/NEJMoa032423
- Tang ASL, Wells GA, Talajic M, et al; Resynchronization-Defibrillation for Ambulatory Heart Failure Trial Investigators. Cardiac-resynchronization therapy for mild-to-moderate heart failure. *N Engl J Med.* 2010;363(25):2385-2395. doi:10.1056/NEJMoa1009540
- Køber L, Thune JJ, Nielsen JC, et al; DANISH Investigators. Defibrillator implantation in patients with nonischemic systolic heart failure. *N Engl J Med.* 2016;375(13):1221-1230. doi:10.1056/NEJMoa1608029
- Moss AJ, Hall WJ, Cannom DS, et al; MADIT-CRT Trial Investigators. Cardiac-resynchronization therapy for the prevention of heart-failure events. *N Engl J Med.* 2009;361(14):1329-1338. doi:10.1056/NEJMoa0906431
- Goldenberg I, Kutiyafa V, Klein HU, et al. Survival with cardiac-resynchronization therapy in mild heart failure. *N Engl J Med.* 2014;370(18):1694-1701. doi:10.1056/NEJMoa1401426
- Gill EA, Poole JE. Will the real left bundle branch block please stand up? *J Am Coll Cardiol.* 2015;66(6):642-644. doi:10.1016/j.jacc.2015.06.1299
- Poole JE, Singh JP, Birgersdotter-Green U. QRS duration or QRS morphology: what really matters in cardiac resynchronization therapy? *J Am Coll Cardiol.* 2016;67(9):1104-1117. doi:10.1016/j.jacc.2015.12.039
- Deshmukh P, Casavant DA, Romanyshyn M, Anderson K. Permanent, direct His-bundle pacing: a novel approach to cardiac pacing in patients with normal His-Purkinje activation. *Circulation.* 2000;101(8):869-877. doi:10.1161/01.cir.101.8.869
- Huang W, Su L, Wu S, et al. A novel pacing strategy with low and stable output: pacing the left bundle branch immediately beyond the conduction block. *Can J Cardiol.* 2017;33(12):1736.e1-1736.e3. doi:10.1016/j.cjca.2017.09.013
- Vijayaraman P, Sharma PS, Cano Ó, et al. Comparison of left bundle branch area pacing and biventricular pacing in candidates for resynchronization therapy. *J Am Coll Cardiol.* 2023;82(3):228-241. doi:10.1016/j.jacc.2023.05.006
- Chung MK, Patton KK, Lau CP, et al. 2023 HRS/APHRS/LAHS guideline on cardiac physiologic pacing for the avoidance and mitigation of heart failure. *Heart Rhythm.* 2023;20(9):e17-e91. doi:10.1016/j.hrthm.2023.03.1538
- Diaz JC, Sauer WH, Duque M, et al. Left bundle branch area pacing versus biventricular pacing as initial strategy for cardiac resynchronization. *JACC Clin Electrophysiol.* 2023;9(8 Pt 2):1568-1581. doi:10.1016/j.jacep.2023.04.015
- Wang Y, Zhu H, Hou X, et al; LBBP-RESYNC Investigators. Randomized trial of left bundle branch vs biventricular pacing for cardiac resynchronization therapy. *J Am Coll Cardiol.* 2022;80(13):1205-1216. doi:10.1016/j.jacc.2022.07.019
- Herweg B, Sharma PS, Cano Ó, et al. Arrhythmic risk in biventricular pacing compared with left bundle branch area pacing: results from the I-CLAS study. *Circulation.* 2024;149(5):379-390. doi:10.1161/CIRCULATIONAHA.123.067465
- Kim JA, Kim SE, Ellenbogen KA, Vijayaraman P, Chelu MG. Clinical outcomes of conduction system pacing versus biventricular pacing for cardiac resynchronization therapy: a systematic review and meta-analysis. *J Cardiovasc Electrophysiol.* 2023;34(8):1718-1729. doi:10.1111/jce.15976